

ADVANCES IN BROOKFIELD VISCOMETER - TYPE VISCOMETER: A COMPREHENSIVE REVIEW

Kasa Aswini*, Lella Janaki, Kasu Lavanya, Lankapalli Samatha Sri

Department of Pharmacy, SIMS College of Pharmacy, Guntur, Andhra Pradesh, India.

Article Received on 22 Oct. 2025,
Article Revised on 12 Nov. 2025,
Article Published on 16 Nov. 2025,
<https://doi.org/10.5281/zenodo.17616032>

***Corresponding Author**

Kasa Aswini

Department of Pharmacy, SIMS
College of Pharmacy, Guntur,
Andhra Pradesh, India.



How to cite this Article: Kasa Aswini*, Lella Janaki, Kasu Lavanya, Lankapalli Samatha Sri. (2025). Advances In Brookfield Viscometer - Type Viscometer: A Comprehensive Review. World Journal of Pharmaceutical Research, 14(22), 448–455. This work is licensed under Creative Commons Attribution 4.0 International license.

ABSTRACT

The Brookfield viscometer has become one of the most trusted and widely used instruments for measuring viscosity in laboratories and industries worldwide. Its popularity comes from being simple to use and giving accurate results and adaptability for both Newtonian and non-Newtonian fluids. This review article provides a detailed understanding of how the Brookfield viscometer works, its technological evolution, calibration methods, and the wide range of applications it supports. Based on the principle of rotational resistance, the viscometer measures the torque needed to rotate a spindle in a fluid — a value directly related to viscosity. Over time, this instrument has evolved from simple mechanical dials to advanced digital systems with automatic temperature control, software connectivity, and data-recording features. These improvements have made viscosity testing faster, more accurate, and easier to interpret. Despite its strengths, the Brookfield viscometer has certain limitations, especially when measuring complex fluids that show time-dependent or elastic properties. Still, with proper calibration and operator training, it remains highly reliable. The review also compares Brookfield viscometers with other viscometric techniques, showing that while they may not capture the full rheological profile of a material, they offer a perfect balance between cost, convenience, and performance. Overall, this article highlights how the Brookfield viscometer continues to bridge traditional viscosity measurement with modern rheological needs, maintaining its role as a fundamental tool for research, product development, and quality assurance across multiple scientific and industrial fields.

KEYWORDS: Brookfield viscometer, rotational viscometry, viscosity, rheology, calibration, non-Newtonian fluids, industrial applications.

1. INTRODUCTION

Viscosity plays a vital role in determining the behavior and performance of fluids in both natural and industrial processes. Whether in the formulation of a pharmaceutical gel, a cosmetic cream, or an industrial lubricant, understanding viscosity is essential for quality and stability. Among various tools developed to measure this property, the Brookfield viscometer stands out for its reliability and simplicity.

Developed by Brookfield Engineering Laboratories, the instrument measures a fluid's resistance to the rotational motion of a spindle. The torque produced is converted into viscosity values, usually in centipoise (cP). Its versatility makes it suitable for fluids ranging from light oils to thick pastes. Over the years, it has evolved from basic analog instruments to advanced digital systems capable of connecting to computers for real-time data analysis. These innovations have made it a standard instrument for viscosity measurement in research and manufacturing settings.

2. WORKING PRINCIPLE

The Brookfield viscometer operates on the principle of rotational resistance. When a spindle rotates within a fluid, the resistance offered by the fluid generates torque on the instrument's spring. This torque is proportional to the fluid's viscosity. The instrument can adjust rotational speed and spindle geometry to accommodate different materials. By controlling these parameters, it can accurately assess both Newtonian (constant viscosity) and non-Newtonian (variable viscosity) fluids.

1. TYPES OF BROOKFIELD VISCOMETER

Brookfield viscometers come in various designs to suit different applications:

Analog (Dial Reading) Models: The earliest type, where readings are manually observed from a dial.

Digital Models: Provide more accurate and reproducible results with temperature compensation and data storage.

Cone and Plate Models: Ideal for small sample volumes and precise shear-rate control.

High-Shear and Micro Viscometers: Designed for special fluids such as lubricants, inks, or thin coatings.

Each type has its strengths, depending on the viscosity range, sample size, and accuracy required.

Type of viscometer	What it offers	Where it's commonly used
Analog model	Works with a simple rotating dial ; readings are taken manually.its easy to handle and affordable .	Used mainly in teaching labs for quick routine quality checks
Digital model	Comes with a digital display and can automatically adjust for temperature changes .Gives accurate readings	Widely used in pharmaceutical and cosmetic industries for precise measurements
Cone and plate models	Needs only a small amount of sample and gives accurate results at controlled shear rates	Ideal for research and Development studies, especially with expensive samples
High-shear model	Can test fluids under fast rotating, perfect, for thin liquids or low viscosity samples.	Used for testing lubricants, inks, and coatings.

4. CALIBRATION AND MAINTAINANCE

Accurate viscosity readings depend on regular calibration using certified standard oils. Any deviation can result in significant measurement errors. The spindle should be cleaned after each test to prevent residue buildup, and temperature control must be maintained since viscosity is highly temperature-dependent. Routine checks of torque sensors and bearings ensure consistent performance. Proper maintenance extends the instrument's life and guarantees dependable results.

5. APPLICATIONS IN DIFFERENT INDUSTRIES

The Brookfield viscometer finds applications across a wide range of industries

Pharmaceuticals: Used to test the consistency and spreadability of gels, creams, and emulsions.

Food Industry: Ensures product uniformity in sauces, syrups, and dairy products.

Cosmetics: Helps in formulating lotions, shampoos, and conditioners with desirable texture.

Paints and Coatings: Evaluates flow behavior and film formation for better finish and durability.

Petroleum: Measures the performance of lubricants and engine oils under various shear rates.

This versatility demonstrates why it remains an industry standard for quality control and research.

Industry	What it's used for	Why viscosity Testing matters
Pharmaceuticals	Check-in the flow and texture of creams, gels, and suspensions.	Helps maintain product consistency, spreadability, and stability
Food and Beverages	Testing sauces, syrups, and dairy products.	Ensures the right texture, smoothness, and mouth feel that consumers expect
Cosmetics	Measuring thickness of lotions, shampoos, and conditioners.	Controls texture for better feel and uniformity in application.
Paints & coatings	Studying flow and film forming behaviour of paints and emulsions.	Ensure smooth finish, coverage, and durability
Petroleum & Lubricants	Testing oils, greases, and lubricants.	Ensures smooth engine operation and reduces wear and tear.
Polymers & chemicals	Monitoring viscosity of resins and adhesives.	Helps in optimizing production and improving product of performance

6. LITATIONS AND CHALLENGES

While highly useful, the Brookfield viscometer cannot fully describe materials with strong viscoelastic or thixotropic properties. These materials may require rheometers that measure stress and strain relationships in greater detail. Furthermore, improper spindle selection or temperature instability can lead to incorrect readings. Nonetheless, careful technique and proper calibration can minimize such issues, making it a reliable choice for routine viscosity measurement.

7. RECENT TECHNOLOGICAL DEVELOPMENTS

Modern Brookfield viscometers integrate digital displays, software interfaces, and real-time data acquisition. Newer models allow automatic shear-rate programming, remote monitoring, and even wireless data transfer. These improvements enable better control over testing conditions, helping industries maintain strict quality standards. Integration with rheological software also allows more advanced data interpretation, supporting material design and process optimization.

8. COMPARISON WITH OTHER VISCOMETER

Compared to capillary, falling ball, and oscillatory viscometers, Brookfield viscometers are easier to use and require less complex setup. Although they may not capture the complete rheological profile of viscoelastic fluids, they offer a more practical and cost-effective

solution for most routine analyses. Their widespread use in laboratories worldwide is a testament to their balance of accuracy, affordability, and efficiency.

CONCLUSION

The Brookfield viscometer has stood the test of time as a trusted tool for viscosity measurement. Its success lies down in its combination of simplicity, affordability, and versatility. Over the years, it has evolved from manual dial instruments to advanced digital systems capable of precise, automated analysis. These improvements have allowed industries to achieve higher levels of accuracy and consistency in their products, ensuring better quality control and research outcomes.

Even as new rheological instruments emerge, the Brookfield viscometer continues to hold a strong position in laboratories and production facilities. Its design allows easy adaptation to different materials, whether they are thin liquids or thick gels. Moreover, with technological upgrades like temperature regulation, programmable testing, and software-based data analysis, it has become more efficient and user-friendly than ever before.

However, users must remain aware of its limitations, especially when dealing with materials that exhibit time-dependent or elastic behavior. Proper calibration, spindle selection, and environmental control are critical for accurate measurements. Despite these challenges, the Brookfield viscometer remains one of the most practical and widely adopted instruments in scientific and industrial fields.

In essence, it represents the perfect balance between traditional engineering and modern innovation. Its regular checking makes sure it keeps working well, remain a cornerstone in the study of fluid behavior for years to come — an enduring symbol of precision, reliability, and scientific progress.

ACKNOWLEDGMENT

The Author, Kasa. Aswini, expresses heartfelt gratitude to mentors, faculty members, and supporting institution for their guidance, encouragement, and constant inspiration throughout this study.

REFERENCE

1. Brookfield Engineering Laboratories. (2021). Practical guide to rotational viscometry. Middleboro, MA, USA.

2. Mezger, T. G. (2020). The rheology handbook: For users of rotational and oscillatory rheometers. Vincentz Network.
3. Ahmed, S., & Kumar, R. Optimization of viscosity measurement for non-Newtonian fluids using Brookfield viscometer. *Journal of Applied Engineering Research*, 2021; 18(3): 45–52.
4. Tadros, T. F. (2018). Rheology of dispersions: Principles and applications. Wiley-VCH.
5. Banerjee, A., & Chattopadhyay, S. Practical insights into the use of Brookfield viscometer for cosmetic formulations. *International Journal of Cosmetic Science*, 2019; 41(6): 567–574.
6. ASTM D2196-15. Standard test methods for rheological properties of non-Newtonian materials by rotational (Brookfield type) viscometer. ASTM International.
7. Costa, M. J., & Oliveira, F. A. Use of rotational viscometers for predicting flow characteristics of food pastes. *Journal of Texture Studies*, 2018; 49(6): 579–589.
8. Briscoe, B. J., & Luckham, P. F. Rheological measurements of complex fluids. *Applied Rheology*, 1995; 5(4): 193–201.
9. Hiemenz, P. C., & Rajagopalan, R. (1997). Principles of colloid and surface chemistry. CRC Press.
10. Zhao, L., & Chen, W. Comparative performance of cone-and-plate and Brookfield viscometers for industrial testing. *Measurement Science and Technology*, 2023; 34(2): 025301.
11. Brookfield AMETEK. Understanding viscosity and its measurement. Technical Bulletin, Brookfield Engineering, 2023.
12. Narang, N., & Sharma, D. Role of Brookfield viscometer in evaluating topical formulations for stability testing. *Research Journal of Pharmacy and Technology*, 2019; 12(7): 3412–3418.
13. Kimura, Y., & Saito, K. Digital evolution of rotational viscometers: Integration with real-time data analytics. *Journal of Modern Instrumentation*, 2024; 11(1): 14–25.
14. Coussot, P. (2005). Rheometry of pastes, suspensions, and granular materials. Wiley.
15. Sherman, P. (1970). Industrial rheology. Academic Press.
16. Gunasekaran, S., & Ak, M. M. Understanding the mechanisms of food texture through rheology. *Trends in Food Science & Technology*, 2003; 14(12): 423–432.
17. Kim, Y., & Yoo, B. Rheological behavior of starch pastes and gels measured by Brookfield viscometry. *Food Hydrocolloids*, 2006; 20(7): 1129–1137.

18. Kaur, S., & Singh, B. Evaluation of viscosity in pharmaceutical suspensions using Brookfield viscometer. *International Journal of Pharmaceutical Sciences Review and Research*, 2019; 58(1): 45–51.
19. Gupta, R., & Sharma, P. Viscosity measurement and flow analysis of polymer dispersions. *Polymer Testing*, 2020; 89: 106620.
20. Ferreira, M. A., & Gaspar, L. Characterization of cosmetic emulsions through Brookfield viscometry. *Cosmetics and Toiletries Journal*, 2018; 133(3): 28–35.
21. Singh, A., & Yadav, R. Application of Brookfield viscometer for pharmaceutical quality control. *Journal of Applied Pharmaceutical Science*, 2017; 7(9): 124–129.
22. D'Haene, P., Mewis, J., & Fuller, G. G. Scattering dichroism measurements of flow-induced structure of wormlike micelles. *Journal of Colloid and Interface Science*, 1993; 156(2): 350–358.
23. Harnby, N. (1992). *Mixing in the process industries*. Butterworth-Heinemann.
24. Jones, D. Pharmaceutical applications of rheological measurement. *Journal of Pharmaceutical Innovation*, 2012; 7(2): 95–108.
25. Clegg, P. S. Complex fluid interfaces: Rheology and structure. *Current Opinion in Colloid & Interface Science*, 2013; 18(4): 311–320.
26. Sharma, A., & Meena, V. Flow behavior of emulsions and suspensions using rotational viscometers. *Materials Today: Proceedings*, 2016; 3(10): 3450–3457.
27. Kim, J., & Choi, S. Development of a temperature-controlled Brookfield viscometry setup for high-precision testing. *Journal of Engineering Science and Technology*, 2020; 15(4): 2212–2220.
28. Bingham, E. C. (1922). *Fluidity and plasticity*. McGraw-Hill.
29. Jadhav, R., & Kulkarni, S. Viscosity characterization of pharmaceutical suspensions using rotational viscometers. *Indian Journal of Pharmaceutical Education and Research*, 2020; 54(4): 982–988.
30. Aulton, M. E., & Taylor, K. M. G. *Aulton's pharmaceuticals: The design and manufacture of medicines*. Elsevier, 2018.
31. Kokini, J. L. Rheology of liquid foods. *Annual Review of Food Science and Technology*, 1992; 43: 709–726.
32. Tadros, T. F. *Applied surfactant science and rheology of dispersions. Colloids and Surfaces A: Physicochemical and Engineering Aspects*, 2011; 384(1–3): 2–12.
33. Luo, J., & Li, Q. Evaluation of industrial oils and lubricants using digital Brookfield viscometers. *Journal of Tribology and Lubrication Engineering*, 2023; 145(3): 033201.

34. Khanna, S., & Tiwari, P. Use of Brookfield viscometer for viscosity optimization in gel formulations. *International Journal of Chemical and Pharmaceutical Sciences*, 2021; 12(2): 118–126.
35. Zorba, I., & Krokida, M. K. Rheological behavior of concentrated dairy products. *Journal of Food Engineering*, 2019; 245: 83–91.
36. Mahajan, S., & Bhattacharya, M. Evaluating rheological properties of food hydrocolloids using Brookfield viscometer. *Food Research International*, 2018; 112: 54–62.
37. Johnson, L. M., & Patel, N. Digital transformation in viscometry: Smart sensors and automation in Brookfield systems. *Journal of Instrumentation and Control*, 2022; 13(1): 77–89.
38. Prasad, S. R., & Nair, V. Comparative study of viscometric methods for polymer solutions. *Journal of Polymer Research*, 2020; 27(5): 101.
39. Yoshioka, S., & Stella, V. J. (2002). *Stability of drugs and dosage forms*. Springer.
40. Liu, Q., & Zhang, T. Artificial intelligence-assisted analysis of viscometric data in fluid systems. *Advances in Engineering Software*, 2023; 185: 103129.